

335-6-5-.23 Requirements for Dental Dischargers

(1) **Applicability.**

(a) Except as provided in subparagraphs (c), (d), and (e) of this paragraph, this rule applies to dental dischargers as defined in paragraph (2) of this rule.

(b) Dental dischargers subject to this rule are not Significant Industrial Users as defined in paragraph 335-6-5-.02(oo), and are not “Categorical Industrial Users” or “industrial users subject to categorical pretreatment standards” as those terms and variations are used in this chapter, as a result of applicability of this rule.

(c) This rule does not apply to dental dischargers that exclusively practice one or more of the following dental specialties: Oral pathology, oral and maxillofacial radiology, oral and maxillofacial surgery, orthodontics, periodontics, or prosthodontics.

(d) This rule does not apply to wastewater discharges from a mobile unit operated by a dental discharger.

(e) This rule does not apply to dental dischargers that do not discharge any amalgam process wastewater to a POTW, such as dental dischargers that collect all dental amalgam process wastewater for transfer to a Centralized Waste Treatment facility as defined in 40 CFR 437.

(f) Dental dischargers that do not place dental amalgam, and do not remove amalgam except in limited emergency or unplanned, unanticipated circumstances, and that certify such to the Department as required in paragraph (5) of this rule are exempt from any further requirements of this rule.

(g) Dental dischargers subject to this rule are not subject to the requirement to obtain an SID permit under rule 335-6-5-.04.

(2) **Definitions.** Whenever used in this rule, the following shall mean:

(a) “Amalgam process wastewater” means any wastewater generated and discharged by a dental discharger through the practice of dentistry that may contain dental amalgam

(b) “Amalgam separator” means a collection device designed to capture and remove dental amalgam from the amalgam process wastewater of a dental facility

(c) “Dental amalgam” means an alloy of elemental mercury and other metal(s) that is used in the practice of dentistry.

(d) “Dental Discharger” means a facility where the practice of dentistry is performed, including, but not limited to, institutions, permanent or temporary

offices, clinics, home offices, and facilities owned and operated by Federal, state or local governments, that discharges wastewater to a publicly owned treatment works (POTW).

(e) “Duly Authorized Representative” as defined in 40 CFR §403.12(l)(3)

(f) “Existing sources” means a dental discharger that is not a new source

(g) “Mobile unit” means a specialized mobile self-contained van, trailer, or equipment used in providing dentistry services at multiple locations.

(h) “New sources” means a dental discharge whose first discharge to a POTW occurs after July 14, 2017.

(3) Compliance Dates for Pretreatment Standards.

(a) No later than July 14, 2020, any existing source subject to this rule must comply with the requirements of paragraphs (4) of this rule unless exempted by paragraph (1)(f) of this rule.

(b) As of July 14, 2017, any new source subject to this rule must comply with the requirements of paragraphs (4) of this rule unless exempted by paragraph (1)(f) of this rule.

(4) Pretreatment Standards.

(a) Removal of dental amalgam solids from all amalgam process wastewater by one of the following methods:

1. Installation, operation, and maintenance of one or more amalgam separators that meet the following requirements:

(i) Compliant with either the American National Standards Institute (ANSI) American National Standard/American Dental Association (ADA) Specification 108 for Amalgam Separators (2009) with Technical Addendum (2011) or the International Organization for Standardization (ISO) 11143 Standard (2008) or subsequent versions so long as that version requires amalgam separators to achieve at least a 95% removal efficiency. Compliance must be assessed by an accredited testing laboratory under ANSI’s accreditation program for product certification or a testing laboratory that is a signatory to the International Laboratory Accreditation Cooperation’s Mutual Recognition Arrangement. The testing laboratory’s scope of accreditation must include ANSI/ADA 108-2009 or ISO 11143.

(ii) The amalgam separator(s) must be sized to accommodate the maximum discharge rate of amalgam process wastewater.

(iii) A dental discharger subject to this rule that operates an amalgam separator that was installed at a dental facility prior to June 14, 2017, satisfies

the requirements of subparagraphs (4)(a)1.(i) and (ii) of this rule until the existing separator is replaced as described in subparagraphs (4)(a)1.(v) of this rule or until June 14, 2027, whichever is sooner.

(iv) The amalgam separator(s) must be inspected in accordance with the manufacturer's operating manual to ensure proper operation and maintenance of the separator(s) and to confirm that all amalgam process wastewater is flowing through the amalgam retaining portion of the amalgam separator(s).

(v) In the event that an amalgam separator is not functioning properly, the amalgam separator must be repaired consistent with manufacturer instructions or replaced with a unit that meets the requirements of subparagraphs (4)(a)1.(i) and (ii) of this rule as soon as possible, but no later than 10 business days after the malfunction is discovered by the dental discharger, or an agent or representative of the dental discharger.

(vi) The amalgam retaining units must be replaced in accordance with the manufacturer's schedule as specified in the manufacturer's operating manual or when the amalgam retaining unit has reached the maximum level, as specified by the manufacturer in the operating manual, at which the amalgam separator can perform to the specified efficiency, whichever comes first.

2. Installation, operation, and maintenance of one or more amalgam removal device(s) other than an amalgam separator. The amalgam removal device must meet the following requirements:

(i) Removal efficiency of at least 95 percent of the mass of solids from all amalgam process wastewater. The removal efficiency must be calculated in grams recorded to three decimal places, on a dry weight basis. The removal efficiency must be demonstrated at the maximum water flow rate through the device as established by the device manufacturer's instructions for use.

(ii) The removal efficiency must be determined using the average performance of three samples. The removal efficiency must be demonstrated using a test sample of dental amalgam that meets the following particle size distribution specifications: 60 percent by mass of particles that pass through a 3150 μm sieve but which do not pass through a 500 μm sieve, 10 percent by mass of particles that pass through a 500 μm sieve but which do not pass through a 100 μm sieve, and 30 percent by mass of particles that pass through a 100 μm sieve. Each of these three specified particle size distributions must contain a representative distribution of particle sizes.

(iii) The device(s) must be sized to accommodate the maximum discharge rate of amalgam process wastewater.

(iv) The device(s) must be accompanied by the manufacturer's manual providing instruction for use including the frequency for inspection and collecting container replacement such that the unit is replaced once it has reached the maximum filling level at which the device can perform to the specified efficiency.

(v) The device(s) must be inspected in accordance with the manufacturer's operating manual to ensure proper operation and maintenance, including confirmation that amalgam process wastewater is flowing through the amalgam separating portion of the device(s).

(vi) In the event that a device is not functioning properly, it must be repaired consistent with manufacturer instructions or replaced with a unit that meets the requirements of subparagraphs (4)(a)2.(i) through (iii) of this rule as soon as possible, but no later than 10 business days after the malfunction is discovered by the dental discharger, or an agent or representative of the dental discharger.

(vii) The amalgam retaining unit(s) of the device(s) must be replaced as specified in the manufacturer's operating manual, or when the collecting container has reached the maximum filling level, as specified by the manufacturer in the operating manual, at which the amalgam separator can perform to the specified efficiency, whichever comes first.

(viii) The demonstration of the device(s) under subparagraphs (4)(a)2.(i) through (iii) of this rule must be documented in the One-Time Compliance Report.

(b) Implementation of the following best management practices (BMP):

1. Waste amalgam including, but not limited to, dental amalgam from chair-side traps, screens, vacuum pump filters, dental tools, cuspidors, or collection devices, must not be discharged to a POTW.

2. Dental unit water lines, chair-side traps, and vacuum lines that discharge amalgam process wastewater to a POTW must not be cleaned with oxidizing or acidic cleaners, including but not limited to bleach, chlorine, iodine and peroxide that have a pH lower than 6 or greater than 8.

(5) Reporting and Recordkeeping Requirements.

(a) Dental Dischargers subject to this rule must comply with the following reporting requirements in lieu of otherwise applicable reporting requirements in this chapter:

1. **One-Time Compliance Report deadlines.** For existing sources, a One-Time Compliance Report must be submitted to the Department no later than October 12, 2020, or 90 days after transfer of ownership. For new sources, a One-Time Compliance Report must be submitted to the Department no later than 90 days following the introduction of wastewater into a POTW.

2. **Signature and certification.** The One-Time Compliance Report must be signed and certified by a responsible corporate officer, a general partner or proprietor if the dental discharger is a partnership or sole proprietorship, or a duly authorized representative in accordance with the requirements of rule 335-6-5-.14.

3. **Contents.**

(i) The One-Time Compliance Report for dental dischargers subject to this rule that do not place or remove amalgam as described in paragraph (1)(f) of this rule must include the following: facility name, physical address, mailing address, contact information, name of the operator(s) and owner(s); and a certification statement that the dental discharger does not place dental amalgam and does not remove amalgam except in limited circumstances.

(ii) The One-Time Compliance Report for dental discharges subject to the pretreatment standards of paragraph (4) of this rule must include:

(I) The facility name, physical address, mailing address, and contact information.

(II) Name(s) of the operator(s) and owner(s)

(III) A description of the operation at the dental facility including: The total number of chairs, the total number of chairs at which dental amalgam may be present in the resulting wastewater, and a description of any existing amalgam separator(s) or equivalent device(s) currently operated to include, at a minimum, the make, model, year of installation.

(IV) Certification that the amalgam separator(s) or equivalent device is designed and will be operated and maintained to meet the requirements specified in subparagraph (4)(a) of this rule.

(V) Certification that the dental discharger is implementing BMPs specified in subparagraph (4)(b) of this rule and will continue to do so.

(VI) The name of the third-party service provider that maintains the amalgam separator(s) or equivalent device(s) operated at the dental office, if applicable. Otherwise, a brief description of the practices employed by the facility to ensure proper operation and maintenance in accordance with paragraph (4) of this rule.

4. **Transfer of ownership notification.** If a dental discharger transfers ownership of the facility, the new owner must submit a new One-Time Compliance Report to the Department no later than 90 days after the transfer.

5. **Retention Period.** As long as a dental discharger subject to this rule is in operation, or until ownership is transferred, the dental discharger or an agent or representative of the dental discharger must maintain the One-Time Compliance Report required at subparagraph (5)(a) of this rule and make it available for inspection in either physical or electronic form.

(b) Dental dischargers or an agent or representative of the dental discharger must maintain and make available for inspection in either physical or electronic form, for a minimum of 3 years:

1. Documentation of the date, person(s) conducting the inspection, and results of each inspection of the amalgam separator(s) or equivalent device(s), and a summary of follow-up actions, if needed.

2. Documentation of amalgam retaining container or equivalent container replacement (including the date, as applicable).

3. Documentation of all dates that collected dental amalgam is picked up or shipped for proper disposal in accordance with subparagraph 335-14-3-.01(4)(a)5., and the name of the permitted or licensed treatment, storage or disposal facility receiving the amalgam retaining containers.

4. Documentation of any repair or replacement of an amalgam separator or equivalent device, including the date, person(s) making the repair or replacement, and a description of the repair or replacement (including make and model).

5. Dental dischargers or an agent or representative of the dental discharger must maintain and make available for inspection in either physical or electronic form the manufacturing operating manual for the current device.

Author: Scott Ramsey.

Statutory Authority: Code of Alabama 1975, §§ 22-22-9, 22-22A-5, 22-22A-6, 22-22A-8.

History: Proposed: June 19, 2020.